
DIGITAL INSTRUMENT

Vibration Meter





Vibration Standard

ISO/IS2373 Motor Quality Standard According As Vibration Velocity					
Quality Rank	Rev (rpm)	H: high of shaft (mm) Maximum vibration velocity rms (mm/s)			
		80 <h<132< td=""><td>132<h<225< td=""><td>225<h<400< td=""></h<400<></td></h<225<></td></h<132<>	132 <h<225< td=""><td>225<h<400< td=""></h<400<></td></h<225<>	225 <h<400< td=""></h<400<>	
Normal	600~3600	1.8	2.8	4.5	
Good (R)	600~1800	0.71	1. 12	1.8	
	1800~3600	1. 12	1.8	2.8	
Excellent (S)	600~1800	0. 45	0.71	1.12	
	1800~3600	0.71	1. 12	1.8	

Model: VM-6380 (3D Vibration Meter) VM-6370 (Enhanced Type)

Applications

Used for measuring periodic motion, to check the imbalance and deflecting of moving machinery. Specifically designed for present measuring various mechanical vibration. So as to provide the data for the quality control, run time and equipment upkeep.

- * VM-6380 can shown 3 same parameters in one display for 3 dimensional measurement.
- * VM-6370 can display the parameters of Displacement, Velocity and Acceleration simultaneously.

Features

- * With the detection of 3 Dimensions, 3D Vibration Meter VM-6380 is the most scientific, comprehensive Vibration Meter in vibration detection field.
- * In accordance with ISO 2954, used for periodic measurements, to detect out-of-balance, misalignment and other mechanical faults in rotating machines.
- * Specially designed for easy on site vibration measurement of all rotating machinery for quality control, commissioning, and predictive maintenance purposes.
- * Individual high quality accelerometer for accurate and repeatable measurements.
- * Wide frequency range (10Hz~10kHz) in acceleration mode.
- * Optional headphones for use as electronic stethoscope.
- * Use RS-232 data output to connect with PC.
- * Provide Bluetooth data output choice.



ติดต่อบริษัท นี โอนิคส์ จำกัด Tel: 098-479-5684 หรือ 061-8268939 E-mail: sale@neonics.co.th เว็บไซต์ www.neonics.biz Specifications

lel	VM-6380	VM-6370		
		Piezoelectric Transducer		
	0.1~400 m/s ² 0.3~1312 ft/s ² 0.0~40 g Equivalent Peak 0.01~400 mm/s 0.04~16.0 inch/s True RMS			
1	0.001~4.0 mm 0.04~160.0 mil Equivalent Peak-peak			
Acceleration	10Hz~10kHz			
Velocity	10Hz~1kHz			
Displacement	10Hz~1kHz			
racy	5% of Reading + 2 digits			
Temperature	0~50 °C			
Humidity	<90 %RH			
Supply	2x1.5V AA (UM-3) Battery			
sions	130x76x32mm			
ght	340 g (Not Including Batteries)			
ccessories	Main Unit			
	3-Axis Piezoelectric Accelerometer	Piezoelectric Transducer		
	Powerful Magnetic Base			
	Probe (Cone) & Probe (Spherical)			
	Carrying Case (B04)			
	Manual Book			
ccessories	Headset			
	RS-232C Data Cable with Software			
	Bluetooth Data Adapter with Software			
	Displacement acy Temperature Humidity supply sions ght ccessories	or3-Axis Piezoelectric AccelerometerAcceleration0.1~400 m/s² 0.3~1312 ft/s²Velocity0.01~400 m/s² 0.3~1312 ft/s²Displacement0.001~4.0 mm 0.04~160.0Acceleration10Hz~Velocity10Hz~Displacement10Hz~Displacement10Hz~accy5% of ReadiTemperature0~55Humidity<90 G		

Accessories

Accessories	Diagram	Using Situations	Using Method
3-Axis Piezoelectric Accelerometer	* * *	Three dimensional vibration parameters measurement of objects.	Be used with Powerful Rare Earth Magnet & Stinger Probe.
Piezoelectric Transducer		General vibration parameters measurement of objects.	Be used with Powerful Rare Earth Magnet & Stinger Probe.
Rare Earth Magnet		Magnetic objects with flat surface, roughness of less than Ra1.6, acceleration ≤ 20 m/s.	connect the vibration sensor with Rare Earth Magnet with the M5 bolt included. And then place the Rare Earth Magnet to the object to be tested.
Stinger Probe (Ball / Cone)	=}	Frequency is less than 1KHz and vibration energy is not small.	Connect the needle to the sensor directly by using probe groupware.